(S8161) 8TRL

S8161



Test Results

ISO/FDIS 7176/19 12th March 2003

Section 5.2- Frontal Impact Performance Requirements

	elchair: 8TRL New Style Frame (18kg) upant: Hybrid II 50 th Percentile (75kg)	RESULTS	
5.2.1	During the Test	[B:	
a).	Horizontal ATD and wheelchair excursion limits as per limits taken from Table 1:-		
Xwc	Was the horizontal movement of the test wheelchair P- Point less than 200 mm. (±5mm)	PASS 68mm	
Xknee	375 mm. (±5mm)	PASS 246mm	
Xhead	ff Was the horizontal movement of the dummy head less than 650 mm. (±5mm)	PASS 421mm	
Xhead	dr Was the horizontal movement of the dummy head less than 400 mm. (±5mm)	Not Measured	
b).	Was the ratio X knee/X wc >1.1	PASS 3.6:1	
c).	Not Measured '		
d).	Did the batteries of powered wheelchairs, or their surrogate parts:-		
	i). move outside of the wheelchair footprint	PASS	
	ii). move into the wheelchair user's space	PASS	
5.2.2	Post Test		
a).	Did the wheelchair remain upright on the test platform	PASS	
	Did the ATD remain in a seated posture in the test wheelchair with a torso angle less that 45° to the vertical	PASS	
b).	Did the wheelchair securement points show visible signs of material failure	PASS	
c).	Did any components of a mass greater that 100g become detached from the wheelchair	PASS	
d).	Did any occupant contactable components fragment or separate with an edge of less than 2mm	PASS	
e).	Did any primary load carrying components of the wheelchair show any visible signs of failure	PASS	
f).	Did any 'tilt in space' locking mechanisms show signs of failure	PASS	
g).	Was the ATD released without the use of tools	PASS	
h).	Was the wheelchair released from the restraint system without the use of tools	PASS	
i).	Was the average decrease of H-Point height relative to	PASS	
	the wheelchair platform less than 20% of the pre-test height.	10%	
Has th	ne wheelchair satisfied the Dynamic Test requirements of DIS 7176/19 of 30 th August 2001	PASS	

Report No.: MBK 03/0150